



2255 116th Avenue N.E., Bellevue, WA 98004-3039, USA

ISO13485

T: 425-455-5200 or 800-888-6616 · F: 425-462-6669 · E: contact@ocularinc.com · I: www.ocularinc.com

Declaration of Conformity

Product Name Sterile Contact Ophthalmic Lens
Model/Number For Detailed list of Model and Numbers, see attached list

Name of Company Ocular Instruments Address of Company 2255 116th Ave NE
Bellevue, WA 98004 USA
Representative Staci Caldwell

Authorized Representative Emergo Europe Address of Authorized Representative Prinsessegracht 20
2514 AP The Hague
The Netherlands
Telephone/email +31.70.345.8570
EmergoEurope@ul.com

Notified Body SGS Belguim NV Notified Body ID Number 1639
CE Certificate Number US19/819943596

Class 1 (Sterile)
Rule 5
Route to Compliance Annex VII, Annex V (Sterility Only) of MDD 93/42/EEC Council Directive

Ocular Instruments declares that the above mentioned products meet the provision of the Council Directive 93/42/EEC for Medical Devices.

Representative Staci Caldwell
Title Quality Manager
Date 01-Feb-21

Signature:

Declaration of Conformity

DECLARATION OF CONFORMITY PRODUCTS

PRODUCT CODE	LONG DESCRIPTION	STOCK CODE	GMDN
OCULUS 55110	Oculus Disposable Flat Vitrectomy, 10 PK	11750	58070
ODV3P	Disposable Vitrectomy Lens 30°, 10 PK	9001	58070
ODVB	Disposable Vitrectomy Lens Bicon, 10 PK	9003	58070
ODVF	Disposable Vitrectomy Lens Flat, 10 PK	8998	58070
ODVM	Disposable Vitrectomy Lens Mag., 10 PK	9000	58070
ODVW	Disposable Vitrectomy Lens Wide, 10 PK	9002	58070
OSIG	Disposable Secureflex HF Surgical Gonio, 10 PK	11813	11894